


DrugControl[®]**BUPRENORPHINE
008A103****ulti
med**TM

Dip the **DrugControl** BUPRENORPHINE TEST strip into urine specimen for at least 10-15 seconds

DIRECTIONS FOR USE

- 1 Allow the test strip, urine specimen, and / or controls to reach room temperature (15 – 30 °C) prior to testing.
- 2 Bring the pouch to room temperature before opening it.
- 3 Remove the test device from the sealed pouch and use it as soon as possible.
- 4 Immerse the test vertically in the urine specimen for at least 10-15 seconds.
- 5 Do not pass the maximum line (dipping line) on the test when immersing the strip.
- 6 Place the test on a non-absorbent flat surface, start the timer and wait for the red line(s) to appear.
- 7 The result should be read at 5-8 minutes. Do not interpret after more than 10 minutes.

INTERPRETATION OF RESULTS

Negative:* Two lines appear. One red line should be in the control region (C), and another apparent red or pink line should be in the test region (T). This negative result indicates that the concentrations of the substances detectable with this test are below the cut-off concentration (substances & cut-off concentrations see table on page 1) or that they are not present.

Positive: One red line appears in the control region (C). No line appears in the test region (T). This positive result indicates that the concentration of at least one of the substances detectable with this test exceeds the cut-off concentration (substances & cut-off concentrations see table on page 1).

Invalid: Control line fails to appear. Insufficient specimen volume or incorrect procedural techniques are the most likely reasons for control line failure. Review the procedure and repeat the test using a new test device. If the problem persists, discontinue using the lot immediately and contact distributor / manufacturer.

* **Note:** The shade of red in the test line region (T) may vary, but it should be considered negative whenever there is even a faint pink line.

ulti med Products (Deutschland) GmbH
Reeshoop 1 • 22926 Ahrensburg • Germany
Telefon: +49-4102 – 80090
Fax: +49-4102 – 50082
e-mail: info@ultimed.de

ulti med Products (Belgium) BVBA
Honzebroekstraat 137 • 8800 Roeselare
Phone: +32 +51 200 425
Fax: +32 +51 200 449
e-mail: belgium@ultimed.org



August 2005
/C